

# Erratum to “Regardless of the Puncture Angle, a Tuohy Needle with Sand-Blasted Steel at the Tip of the Bevel Is a Valid Needle on the Ultrasonic Images”, [Open Journal of Anesthesiology, 2016, 6, 101-104]

Ichikawa Yuki<sup>1</sup>, Hironobu Ueshima<sup>2\*</sup>, Akira Kitamura<sup>1</sup>

<sup>1</sup>Department of Anesthesiology, Saitama Medical University International Medical Center, Saitama, Japan

<sup>2</sup>Department of Anesthesiology, Showa University Hospital, Tokyo, Japan

Email: \*ueshimhi@yahoo.co.jp

**How to cite this paper:** Yuki, I., Ueshima, H. and Kitamura, A. (2021) Erratum to “Regardless of the Puncture Angle, a Tuohy Needle with Sand-Blasted Steel at the Tip of the Bevel Is a Valid Needle on the Ultrasonic Images”, [Open Journal of Anesthesiology, 2016, 6, 101-104]. *Open Journal of Anesthesiology*, 11, 219-220.

<https://doi.org/10.4236/ojanes.2021.117021>

**Received:** June 27, 2021

**Accepted:** July 16, 2021

**Published:** July 19, 2021

Copyright © 2021 by author(s) and Scientific Research Publishing Inc.

This work is licensed under the Creative Commons Attribution International License (CC BY 4.0).

<http://creativecommons.org/licenses/by/4.0/>



Open Access

## Abstract

The original online version of this article (Ichikawa Yuki, Ueshima Hironobu, Hiroshi Otake, Akira Kitamura (2016) Regardless of the Puncture Angle, a Tuohy Needle with Sand-Blasted Steel at the Tip of the Bevel Is a Valid Needle on the Ultrasonic Images, 2016, 6, 101-104, [https://file.scirp.org/Html/1-1920418\\_68780.htm](https://file.scirp.org/Html/1-1920418_68780.htm)) unfortunately contains some mistakes. The authors wish to correct the errors in co-authors, Table 1, and Table 2.

Our article had some mistakes in co-authors, **Table 1**, and **Table 2**.

The followings are the list of errata and correct tables

## List of Errata

	TRUE	FALSE
P.101 Co-Authors	Yuki Ichikawa, Hironobu Ueshima, Akira Kitamura	Yuki Ichikawa, Hironobu Ueshima, Hiroshi Otake, Akira Kitamura

## Table 1

	the depth from Blue Phantom (cm)			
	0.5	1.0	1.5	2.0
Rafa	193.7 (6.5)	183.3 (6.0)	189.3 (9.1)	182.1 (5.7)
Non-coated	133.8 (3.8)	116.9 (8.2)	94.2 (4.2)	67.3 (1.5)

**Table 2**

	the depth from Blue Phantom (cm)					
	0.5	1.0	1.5	2.0	2.5	3.0
Rafa	194.0 (7.3)	188.3 (5.7)	187.2 (4.9)	179.4 (4.9)	172.0 (4.5)	163.8 (7.5)
Non-coated	134.4 (2.7)	111.6 (7.3)	104.5 (5.8)	86.1 (7.2)	72.2 (3.9)	56.6 (4.6)