

The Basis for Design and Manufacture of a DSP-Based Coincidence Spectrometer

Nguyen Xuan Hai^{1*}, Pham Ngoc Tuan¹, Nguyen Nhi Dien¹, Dang Lanh¹, Tuong Thi Thu Huong¹,
Pham Dinh Khang²

¹Nuclear Research Institute, Dalat, Vietnam; ²Nuclear Training Center, Hanoi, Vietnam.
Email: *nxhai@hcm.vnn.vn

Received July 1st, 2013; revised August 8th, 2013; accepted August 17th, 2013

Copyright © 2013 Nguyen Xuan Hai *et al.* This is an open access article distributed under the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly cited.

ABSTRACT

The coincidence technique and the coincidence spectroscopy have been developed and applied for over 40 years. Most of popular coincidence measurement systems were based on analog electronics techniques such as time to amplitude conversion (TAC) or logic selecting coincidence unit. The above-mentioned systems are relatively cumbersome and complicated to use. With the strong growth of digital electronics techniques and computational science, the coincidence measurement systems will be constructed simpler but more efficient with the sake of application. This article presents the design principle and signal processing of a simple two-channel coincidence system by a technique of digital signal processing (DSP) using Field Programmable Gate Arrays (FPGA) devices at Nuclear Research Institute (NRI), Dalat.

Keywords: Coincidence Spectrometer; Digital Signal Processing; FPGA

1. Introduction

The background reduction measurement systems with active methods are based on coincidence or anti-coincidence techniques. They are mainly built from proper functional electronics modules in NIM or CAMAC standards [1,2]. They allow us to identify coincidence or anti-coincidence events via main electronics blocks called coincidence unit or time to amplitude converter. Normally, basic configuration of a coincidence measurement system will, at least, consist of two channels, and the selection of coincidence event pairs depends on the defining moment of the pulses appearing at “Timing” output of the pre-amplifier. Obvious drawback of this system is cumbersome in size, adjusting operation and synchronizing signals among electronics stages.

The growth of computer engineering and programmable devices that are capable of operating at high frequencies has allowed us to design a new spectrometry generation. The spectrometry generation is compact on size, simple in terms of connectivity and using [2-4].

At Nuclear Research Institute (NRI), Dalat, a number of researches and construction of Compton suppression as well as event-event coincidence systems were performed in 1990s. The results of these studies were reported

in [5]. In the recent period at NRI, further studies on the coincidence spectrometer were presented in several publications [6-8]. Overall, although there have been significant improvements on acquiring as well as processing data, this system is still based on the traditional way in obtaining signals under the operation of a coincidence unit or TAC. With the research results gathered during the installation, investigation and exploitation of an “event-event” coincidence spectrometer on digital signal processing (DSP) at NRI, we have now proposed a new approach for a design of the DSP-based multi-application coincidence technique through Field Programmable Gate Arrays (FPGA) devices. The basis for the design of this spectrometer will be presented within the framework of this article.

2. The Basis and Method for the Design

2.1. Design of Multi-Channel Analyzer

The block diagram of MCA is presented in **Figure 1**. The functions of the diagram can be summarized as follows:

The pre-filter works as high-pass filter giving an output signal with a constant shaping time. In addition, the pre-filter has amplification function to generate the appropriate signals for ADC conversion.

The pre-filter’ output signals are sent to flash ADC for

*Corresponding author.




Figure 1. The MCA block diagram.

sampling. The ADC' output signals are digital copies of input' analog signals.

The transferring "Exponential Decay to Rectangular Shape" circuit converts exponential decay to rectangular pulse, and its function is the opposite of high pass deconvolver (HPD). The transferring "Rectangular Shape to Trapezoidal Shape" circuit converts shape of rectangular to trapezoidal pulse, and its function is the low pass filter (LPF). The circuits of pulse pile-up rejection, base line restoration and built-in configuration are also designed in the main board.

The pulse pile-up rejection circuit detects the pile-up pulses in duration from rise time to half of the flat width, in slow channel. In case of non-overlapped pulse is detected in monitoring duration, the right pulse will be analysis. A dual port memory is integrated in FPGA device for buffering data. The preset time is set by user with capacity up to 4.2×10^9 seconds. The interface circuit is integrated in the main board and connected to PC through a USB-RS232 bridge.

2.2. Design of the "Event-Event" Coincidence Spectrometer

The block diagram of the "event-event" coincidence spectrometer is presented in **Figure 2**. According to the block diagram, the coincidence spectrometer consists of three parts; they are two DSP-based multi-channel analyzers and one channel of timing discriminators. Two MCA channels are used for getting the energy spectrum from the two detectors. Timing discrimination circuits are used for the time interval measurement. The pre-filter, Fast ADC and DSP MCA are the components of the energy analysis channel. The FAST COMPARATOR, TIME COUNTER, TIME COMPARATOR and COINCIDENT CONTROLLER are the components of the coincidence channel.

Except for the PRE-FILTER, FAST ADC and DDR2 MEMORY, almost other components are designed and integrated into one FPGA device which is named XC6SLX16.




Figure 2. The block diagram of the "event-event" coincidence spectrometer.

The external DDR2 MEMORY is used to save the measured results.

The interface circuit is integrated on the main board and connected to PC through USB-RS232 bridge.

Principal Operation

Two MCAs receive pulses from the preamplifier's energy outputs, the pulses' amplitudes are analyzed (by DSP techniques); simultaneously, the preamplifier's output pulses are connected to fast comparator circuits. The output signals of these comparator circuits are used to start/stop a time counter. The used clock frequency is 400 MHz (2.5 ns period). The contents of the counter (time interval between two events) are compared with values of preset time interval; if the measured time is in the selected range, the amplitudes of pair of events will be recorded. Three values E1, E2 and T will be stored into memory; in which E1 and E2 are energy values of pair of coincidence gamma rays, and T is time interval. The measured results are read by data acquisition program and saved to file on the hard disk.

3. Fabricating

3.1. Hardware

The entire design is programmed and configured in the FPGA chip XC6LX16-CS324. The used main board is SP601 kit supplied by Xilinx as **Figure 3**.

3.2. Data Acquisition Program

The data acquisition program for energy and timing spectra was written under LabVIEW 8.5. The functions of program are as follows:

- + Connecting peripheral devices to PC and interfacing to PC through USB port.
- + Control of data acquisition for multi-MCA mode: start/stop data acquisition, preset measurement time, data saving...
- + Control of data acquisition for "event-event" coincidence mode: start/stop data acquisition, preset measurement time, data saving...
- + The base data analysis: display spectra, counts per channel, energy calibration, zoom in and zoom out spectra.

3.3. The Main Features of Spectroscopy

- 1) The multi-channel analysis:
 - Input: accept the preamplifier's output pulses from semiconductor detector, the amplitude from tens to 500 mV;
 - Resolution: 8192 channels;
 - Capacity of counts/channel: $2^{32} - 1$;
 - Preset time capacity: $2^{32} - 1$ seconds with 1 second




Figure 3. The main board.

- step;
- The integral nonlinearity: $<\pm 0.05\%$;
 - Shaping time: 0.4; 0.8; 1.6; 3.2; 6.4; 12.6; 25.2 μs ;
 - Coarse gain: 1, 2, 5 and 10;
 - Fine gain: $0.75 \div 1.25$, step 0.01;
 - Displaying and rejecting the pile-up pulses, baseline correction and dead time correction.
- 2) Event-event coincidence:
- Preset coincidence time window: $5 \div 10,000$ ns, resolution step 5 ns;
 - Mode: coincidence, PHA;
 - Memory capacity: 16 M events.
- 3) Power supply:
- +12 V/800 mA,
 - 12 V/50 mA,
 - +6 V/300 mA,
 - 6 V/100 mA.
- 4) Dimensions: width NIM 2M.

4. Testing and Results

The experiment configuration is showed in **Figure 4**. The disk-shaped ^{60}Co source with activity of 24 kBq was placed between the two detectors (GMX35). The distance from source to the detectors' surfaces was 4 cm. In front of each detector, a sheet of lead 2 mm in thickness




Figure 4. Experiment configuration.

was placed to reduce the effect of back scattering from the detectors to each other.

Data were collected during 2 h. The interfacing program was set up in event-event coincidence mode. The data files were saved to hard disk automatically in all 4096 coincidence events. The **Figures 5-7** are energy spectra of each channel, summation spectra and timing spectra respectively.

Figure 6 shows the ^{60}Co summation spectra, with only one peak and Compton scattering background. The coincident spectrum (**Figure 5**) of each channel contains two




Figure 5. Energy spectrum of ^{60}Co .




Figure 6. Summation spectra of ^{60}Co .




Figure 7. Timing spectra (one clock unit equivalent 2.5 ns).

peaks (1173 and 1332 keV) as in MCA mode with gate control. The measured energy resolution of 1332 keV peak was ~ 8.5 keV, which was worse than three times in comparison with detectors' nomination value.

Figure 7 shows the timing spectra measured with ^{60}Co source. The time resolution and the peak/background height ratio were ~ 12.5 ns and 2453:220 respectively. The timing resolution was good but the spectrum shape

was not smooth.

Figure 6 shows only the full-energy peak of γ -rays of cascade. In practice, the nuclide ^{60}Co decay β -emission to 2.5 MeV level in ^{60}Ni . The 2.5 MeV level is de-excited by a cascade of 1.17 and 1.33 MeV γ -rays to the ground state. None of cross-over is observed. This showed that the data acquisition algorithm and program for DSP coincidence spectrometer are exact but the energy resolution and timing spectra are not good.

5. Conclusions

The design of systems based on DSP techniques using FPGA allows constructing a simple, compact and impact coincidence spectrometer in which all of parameters are selected and controlled by software.

Sample frequency is very important and reduced noise is a big problem. These are problems that should be solved in designing and manufacturing DSP coincidence spectrometer.

Currently, the reduced noise for DSP coincidence spectrometer has been studying at the Department of Nuclear Physics and Electronics, NRI, Dalat. It is hopeful that, in the near future, the design of system might be the basis for development and application of coincidence measurement techniques in the field of physics research and applications.

REFERENCES

- [1] V. T. Jordanov and G. F. Knoll, "Digital Synthesis of Pulse Shape in Real Time for High Resolution Radiation Spectroscopy," *Nuclear Instruments and Methods in Physics Research Section A*, Vol. 345, No. 2, 1994, pp. 337-345.
- [2] M. Lauer, "Digital Signal Processing for Segmented HPGc Detectors Preprocessing Algorithms and Pulse Shape Analysis," Doctor of Science Thesis, University of Heidelberg, Heidelberg, 2004.
- [3] W. R. Leo, "Techniques for Nuclear and Particle Physics Experiments," Springer-Verlag, Berlin, Heidelberg, 1987.
- [4] Valentin T. Jordanov and Glen F. Knoll, "Digital Techniques for Real-Time Pulse Shaping in Radiation Measurements," *Nuclear Instruments and Methods in Physics Research Section A*, Vol. 353, No. 1-3, 1994, pp. 261-264.
- [5] <http://www.fastcomtec.com/products/multiparameter-multichannel-analyzers/mpa-3rtc.html>
- [6] A. Kimura, Y. Toh, M. Koizumi, A. Osa, M. Oshima, J. Goto and M. Igashira, "Development of a Data Acquisition System for a Multiple Gamma-Ray Detection Method," *AIP Conference Proceedings*, Vol. 769, Melville, New York, 2005.
- [7] A. M. Hoogenboom, "A New Method in Gamma-Ray Spectroscopy: A Two Crystal Scintillation Spectrometer with Improved Resolution," *Nuclear Instruments*, Vol. 3,

No. 2, 1958, pp. 57-68.

[doi:10.1016/0369-643X\(58\)90092-6](https://doi.org/10.1016/0369-643X(58)90092-6)

Instruments and Methods in Physics Research A, Vol. 439, No. 2-3, 2000, pp. 378-384.

[doi:10.1016/S0168-9002\(99\)00897-9](https://doi.org/10.1016/S0168-9002(99)00897-9)

- [8] A. Pullia, "Quasi-Optimum Gamma and X-Ray Spectroscopy Based on Real-Time Digital Techniques," *Nuclear*