

Fitting Nonlinear Calibration Curves: No Models Perfect

Julia Martin, Alberto Romero Gracia, Agustín G. Asuero

Department of Analytical Chemistry, The University of Seville, Seville, Spain

Email: asuero@us.es

How to cite this paper: Martin, J., Gracia, A.R. and Asuero, A.G. (2017) Fitting Nonlinear Calibration Curves: No Models Perfect. *Journal of Analytical Sciences, Methods and Instrumentation*, 7, 1-17.
<https://doi.org/10.4236/jasmi.2017.71001>

Received: January 5, 2017

Accepted: February 28, 2017

Published: March 3, 2017

Copyright © 2017 by authors and Scientific Research Publishing Inc.
This work is licensed under the Creative Commons Attribution International License (CC BY 4.0).

<http://creativecommons.org/licenses/by/4.0/>



Open Access

Abstract

The study of the calibration of a series of compounds of environmental concern (six perfluoroalkyl compounds (perfluorooctane sulfonic acid and five perfluoroalkyl carboxylic acids), three preservatives (methyl-, ethyl- and propylparabens) and the brominated flame retardant hexabromocyclododecane) by LC-MS/MS has been carried out, with a view to their simultaneous determination in samples of environmental interest. In some cases nonlinear calibration curves are obtained, but restricting the concentration range a linear model may be used to fit the data. Residual analysis has been performed in order to verify which models fit the data better, opting for a compromise decision given the apparent complexity of residuals plots. As Box states there are no perfect models (but models that work better than others).

Keywords

Calibration, Non Linear Calibration, Residual Analysis, Liquid Chromatography-Tandem Mass Spectrometry, Emerging Pollutants

1. Introduction

Method validation is an important requirement in the practice of chemical analysis. General requirements in method validation for performance characteristics include, but are not limited to, linearity, accuracy, precision, sensibility and robustness [1] [2] [3]. Method validation is, therefore, an essential component of the measures that a laboratory should implement to allow it to produce reliable analytical data. This paper deals on the first ones: Linearity (calibration).

Calibration is an essential part of every quantitative analytical method [3]-[10] and correct performance of the so important step is a critical part of method development and validation.

Calibration is a procedure to standardize the instrument by determining the deviation between a measurement system and a reference system represented by

reference materials and their accepted values. Considering that the majority of analytical methods show linear relationships in one way or another, the recommended statistical methods to be used for the assessment of linearity are ordinary least squares regression or weighted least squares regression [3].

Linearity is described as the ability of the method to elicit test results that are directly proportional to analyte concentration in a given range [5] [6] [7]. In practice, the range is the interval between the upper and lower levels of analyte for the intended analytical method, and for which acceptable precision and accuracy are obtained [3].

However, for some analytical techniques, the relationship between the measured signals and the analyte concentrations is nonlinear and nonlinear or polynomial models are better fitted instead, *i.e.*, a commonly observed phenomenon in atomic absorption spectrophotometry [8] is the ending of the calibration graph towards the concentration axis at elevated concentrations. In most real problems, the response becomes non-linear when the range of the calibration data becomes sufficiently large. In the field of liquid chromatography coupled to tandem mass spectrometry (LC-MS/MS) for instance matrix-related non-linearity can be observed [4] [11] [12] in several methods.

It is well known that when a wrong equation is fitted to data, the shape and the pattern of the residual plot contain valuable information that can be used to determine the way [13]-[20] in which the equation should be modified to achieve a better description of the data. So, residuals provide a convenient means of checking whether the calibration data is actually linear [21] [22] [23] [24]. The residuals are the vertical distances indicated in the y -direction between the points and the regression line (which gives a minimum sum of their squares) [21]. No rigorous mathematical treatment is required. If there is a true linear relationship between the variables with the error symmetrically distributed, residuals will be scattered randomly above and below zero, an equal number of plus and minus. Systematic deviations may indicate either a systematic error in the experiment or an incorrect or inadequate model. A curvilinear pattern in the residuals plot means that a non-linear curve, containing higher order terms, will be better fitted. A linear trend (descending or ascending) may indicate that an additional term in the model is needed. The “fan-shaped” residual pattern shows that experimental error increases with mean response (heteroscedasticity) so the constant variance assumption is inappropriate [21]. This last should be approached by weighted least squares method or by transforming the response.

Among the various statistical ways of numerically measuring some of the observed discrepancies, to date the most widely used method is still the visual examination of the residual plots because it gives more information in a direct way [21]. The simplest model or the model with the minimum number of parameters that adequately fit the data in question is usually the best choice [25]: “*Non sunt multiplicanda entia praeter necessitatem*” (Occam’s razor). However, things as we will have opportunity to see, are not always so simple and so easy.

The use of LC and MS have proved to be a powerful tool for the identification

and quantification of these emerging pollutants in complex mixtures and/or for confirming their presence [26] [27] [28]. In this work, a LC-MS/MS preliminary study on compounds of environmental significance, *i.e.* perfluoroalkyl compounds, preservatives and brominated flame retardants, is carried out taken in mind their simultaneous determination in environmental samples because of their widespread use, potential toxicity, persistence or bioaccumulation [27]-[39]. Calibration curves were obtained and residual analysis [13] [14] has been applied in an attempt to check for model adequacy.

2. Simultaneous Determination of Three Parabens, Six Perfluoroalkyl Compounds and a Flame Retardant Made by LC-MS/MS

Calibration curves are prepared for the simultaneous determination of three parabens (methylparaben (MeP), ethylparaben (EtP), propylparaben (PrP)), six perfluoroalkyl compounds (perfluorooctane sulfonic acid (PFOS), perfluorooctanoic acid (PFOA), perfluoroheptanoic acid (PFHpA), perfluorohexanoic acid (PFHxA), perfluoropentanoic acid (PFPeA), perfluorobutanoic acid (PFBuA)) and a brominated flame retardant hexabromocyclododecane (HBCDD) by LC-MS/MS detection.

2.1. Brominated Flame Retardants

Brominated flame retardants are used in a wide variety of commercial products (furniture, plastics, fabrics, paints, electronic devices) to reduce their flammability [29] [30]. There are currently about 20 - 25 classes of brominated flame retardants, three of which are the main ones: tetrabromobisphenol A and its derivatives, polybrominated diphenyl ethers and hexabromocyclododecane (including three isomers). The concern for these compounds lies basically in their great ubiquity, since they have been detected in a wide range of human, animal and environmental samples [30] [31]. Indications of possible adverse effects [30] [31] [32] such as neurotoxicity, endocrine disruption and cancer have triggered the alarm and the consequent adoption of legislative measures for its control in water at European level [33].

2.2. Perfluoroalkyl Compounds

Perfluorinated detergents are compounds for industrial use in a wide range of sectors. They are nowadays recognized as very dangerous pollutants and they are widely dispersed in the environment [34]. At the heart of the controversy are PFOS and PFOA. PFOS has been used as coolant, detergent, water and oil repellents, flame retardants, lubricants, adhesives, cosmetics, insecticides, etc. PFOA, on the other hand, is used in the manufacture of fluoropolymers (PTFE) and fluoroelastomers (PVDF) and also found used as fabrics, carpets, food containers, automobiles manufacture, etc. Both PFOA and PFOS compounds, according to recent studies, are toxic and persistent [31] [35] [36], PFOA is also carcinogenic, and PFOS has a strong tendency to bioaccumulate.

2.3. Personal Care Products or Preservatives

Parabens (methyl, ethyl, propyl, benzyl, butyl parabens) are among the most commonly used synthetic preservatives in personal care cosmetics and pharmaceuticals, given their supposed low toxicity, broad spectrum of activity, inertia, widespread acceptance in international regulations, biodegradability and low cost [37]. But currently there is a tendency to avoid the use of these compounds due to increasing evidence of its effects of altering the endocrine system [38] [39].

2.4. Materials and Methods

Reagents. The compounds studied (the parabens MeP, EtP and PrP, the perfluoroalkyl compounds PFOS, PFOA, PFHpA, PFHxA, PFPeA and PFBuA and the brominated flame retardant HBCDD) were supplied by Dr. Ehrenstorfer GmbH (between 97% - 99.5% purity). The stock solution of each compound (1000 mg/L) was prepared in methanol and stored in a refrigerator at 4°C. Working solutions were prepared by diluting the stock standard solutions in methanol. Acetonitrile, water and methanol all of HPLC quality purity were supplied by Romil Ltd. (Barcelona, Spain). Ammonium acetate (reagent grade analysis) was supplied by Panreac (Barcelona, Spain).

Liquid chromatography and detector. High-resolution liquid chromatography (Agilent Series 1200, Agilent Technologies, Santa Clara, CA) equipped with a vacuum degasser, a binary pump, an autosampler and a thermostated column compartment (**Figure 1**). Zorbax Eclipse XDB C-18 Rapid Resolution column (50 × 4.6 mm i.d., 1.8 µm). Precolumn XDB C-18 (4 × 4 mm, 5 µm). Mass spectrometry detector (Agilent 6410 Series, Agilent Technologies, Santa Clara, CA) triple quadrupole (QqQ-MS) equipped with electrospray ionization source (ESI).

2.5. Chromatographic Analysis

Analytes were separated using an HPLC system equipped with XDB-C 18 column reverse column of 4.6 mm × 50 mm and 1.8 µm particle size (Agilent Technologies, Santa Clara, CA).



Figure 1. LC-MS/MS equipment.

Our aim was to obtain high sensitivity and selectivity in a short time. First, the pH of mobile phase was studied and deionised water with different additives was studied as aqueous solvent. Acetic acid (from 0% to 0.2%, v/v), ammonia (from 0% to 0.050%, w/v) and mixtures of them (ammonium acetate) were assayed. Higher responses and better peak shapes were obtained using 10 mM ammonium acetate as aqueous solution and methanol as organic solvent. Second, we analyzed the effect of substituting methanol for acetonitrile but no improvements were observed in peak shapes or resolution, so we selected the mobile phase previously mentioned. A linear gradient, as described in **Table 1**, was used. The flow rate was 0.6 mL/min.

Lastly, we increased the injection volume in order to enhance the analytical signal and consequently the limits of detection of the method. A range from 5 to 20 μL was analyzed and 20 μL was chosen as injection volume since a marked increase in sensitivity without loss of resolution was obtained. The increase of temperature from 30°C to 50°C did not improve significantly the characteristics of chromatographic method, therefore 30°C was chosen as optimum.


The HPLC system is coupled to a triple quadrupole mass spectrometer with ESI working in negative mode. The parameters selected for the spectrometer are: capillary voltage, 3000 V; nebulizer pressure, 40 psig; drying-gas flow rate, 9.0 L/min and drying-gas temperature, 355°C. The mode of operation of the spectrometer is MRM (Multiple Reaction Monitoring). Instrument control and data acquisition were carried out with Mass Hunter software (Agilent, USA). A previous optimization of the conditions of fragmentation was made using the Optimizer software. The MS/MS detection method was set up by continuous infusion of standard solutions of each individual compound ($1 \text{ mg}\cdot\text{L}^{-1}$) to optimize the response of the precursor ion. The mass spectrometric conditions were optimized for each compound. ESI interface in positive and negative modes were evaluated. Negative mode was selected because it showed higher sensitivity for all compounds of interest. The two transitions, one for quantification and the other for confirmation, corresponding to the most abundant ion products were selected after the rupture of the precursor ion, in accordance with Decision 2002/657/EC [40]. The most abundant transition ion was selected to obtain maximum sensitivity for quantification. The parameters optimized for product ions were fragmentation voltage and collision energy. The parameters selected to obtain optimum responses are presented in **Table 2**. **Figure 2** and **Figure 3** show the mass spectrum corresponding to MeP and a chromatogram of a standard solution of the compounds under study, at a concentration of 100 ng/mL,

Table 1. Gradient program.

Time (min)	Aqueous phase (% v/v)	Organic phase (% v/v)
0	72	28
20	5	95
20.1	72	28
26	72	28

Table 2. Optimized parameters for the determination of contaminants by QqQ-MS.

Compound	Retention time (min)	Precursor ion (m/z)	MRM 1 (quantification) (m/z)	MRM 2 (confirmation) (m/z)	Fragmentor (V)	Collision energy (V)
MeP	6.225	151.2	92.1	136.1	70	16
EtP	9.006	165.2	92.1	137.1	79	20
PrP	11.578	179.2	92.1	136.1	99	24
PFBuA	3.322	213	169	51.6	55	0
PFPeA	7.302	263	219	89.7	55	0
PFHxA	10.453	313	269	119	60	0
PFHpA	12.647	363.1	319	332.8	65	0
PFOA	14.268	413.1	369.1	194.3	62	0
PFOS	15.595	499	80	51.5	145	40
HBCDD	21.918	640.7	81	79	67	40

**Figure 2.** Mass spectrum of MeP.

obtained after the fragmentation performed under the selected optimum conditions.

2.6. Standards for the Calibration Procedure

Prepare the calibration standards containing concentrations of the compounds in the concentration ranges 1 to 1500 ng/mL. Use methanol as solvent.

2.7. Calibration Curves

They are obtained from the peak areas of MRM (Multiple Reaction Monitoring) chromatograms. The results obtained are shown in **Table 3**, and are shown in **Figures 4-6**.




Figure 3. Chromatogram of a standard solution (100 ng/mL) of the studied compounds.




Figure 4. Response (peak area of MRM chromatograms) versus concentration (calibration curve) obtained by simple linear regression (top) and residual graph (bottom) for MeP, EtP, PrP and PFBuA.




Figure 5. Response (peak area of MRM chromatograms) versus concentration (calibration curve) obtained by simple linear regression (top) and residual graph (bottom) for PFPeA, PFHxA, PFHpA and PFOA.




Figure 6. Response (peak area of MRM chromatograms) versus concentration (calibration curve) obtained by simple linear regression (top) and residual graph (bottom) for PFOS and HBCDD.

3. Results and Discussion

A glance at **Figures 4-6** reveals that the pattern of the residuals obtained by simple linear regression is clearly curvilinear in all cases except for MeP, EtP and HBCDD, which is not surprising given the wide concentrations range used in the calibration process. This is, moreover, typical in instrumental analysis [8] [9] [11] [12], as has been indicated previously. It may also be stressed that the dispersion of the measurements in terms of absolute standard deviation increases with increasing concentration, a circumstance also typical in instrumental analysis, and specifically in LC-MS-MS. **Figure 7** and **Figure 8** plot the standard deviation (SD) and the coefficient of variation ($CV = SD/MEAN$), for the sake of comparison, determined from quadruplicate standard measurements (see **Table 3**) during the same day and at the concentration ranges from 1 to 1500 ng/mL.

This leads us, once found the appropriate model, to the need to apply the weighted least squares method in the calibration process, once the Cochran test shows that the variances are not homogeneous. The coefficient of variation (relative standard deviation) can be considered constant in all the cases, except at low concentrations, in which an increase of the same takes place (**Figure 7** and **Figure 8**). This circumstance is also typical of the instrumental analysis [21]




Figure 7. Standard deviation (SD) and coefficient of variation (CV) as a function of concentration (log scale) for MeP, EtP, PrP, PFBuA, PFPeA and PFHxA.




Figure 8. Standard deviation (SD) and coefficient of variation (CV) as a function of concentration (log scale) for PFHpA, PFOA, PFOS and HBCDD.

Table 3. Experimental data in the LC-QqQ-MS assay of the studied compounds.

Concentration ng/mL	Area									
	MeP	EtP	PrP	PFBuA	PFPeA	PFHxA	PFHpA	PFOA	PFOS	HBCDD
1	1184	1340	2229	6467	9163	13,466	30,167	12,569	2696	132
1	908	1137	2171	7507	9146	14,258	30,377	13,811	2548	105
1	864	1399	2150	7339	9110	12,911	32,345	13,361	2672	167
1	1005	1007	2314	7109	9745	12,663	29,847	12,754	2632	196
5	4492	6073	10728	38,352	40,308	71,226	132,779	67,649	12,570	667
5	4262	5711	9668	37,203	40,531	70,113	132,374	69,896	12,535	593
5	4525	6129	9993	37,667	42,329	71,547	135,051	70,957	12,372	631
5	4271	6212	9642	36,909	41,221	72,233	131,190	70,975	12,756	625
20	15,478	23,380	38,085	147,775	165,675	297,727	541,384	280,281	51,961	2389
20	16,958	23,246	36,226	153,665	171,048	308,450	536,253	292,393	53,230	2279
20	16,593	23,507	37,713	151,629	174,481	303,468	547,111	290,667	52,560	2554
20	16,186	25,234	38,299	153,143	172,763	307,997	556,749	289,058	53,156	2394
75	63,420	94,161	155,639	589,499	679,238	1,173,978	1,970,518	1,124,698	212,557	8529
75	63,347	92,955	153,754	582,509	670,764	1,149,246	1,955,860	1,119,402	209,053	8161
75	64,501	95,575	154,614	583,523	672,129	1,172,311	2,012,733	1,140,991	213,667	8311
75	64,497	97,015	153,771	582,923	679,185	1,183,564	1,984,695	1,128,084	210,118	8618
100	87,532	125,687	209,051	768,774	881,362	1,508,280	2,520,453	1,474,322	279,707	10,660
100	87,654	135,547	211,545	764,762	891,624	1,497,562	2,595,680	1,477,923	284,973	11,273
100	86,192	128,932	210,498	764,999	875,655	1,504,972	2,502,917	1,442,845	279,531	10,894
100	86,281	133,186	208,813	770,867	900,274	1,536,909	2,571,226	1,486,015	281,691	11,179
200	185,653	269,465	442,280	1,451,923	1,657,897	2,789,214	4,520,435	2,735,448	517,434	22,914
200	179,164	270,382	438,705	1,437,231	1,653,882	2,746,647	4,458,679	2,744,527	511,369	21,794
200	180,963	270,409	438,777	1,424,945	1,636,368	2,759,462	4,506,357	2,706,446	511,528	22,728
200	180,617	275,582	441,375	1,432,895	1,651,618	2,749,447	4,559,778	2,739,220	519,854	22,597
400	371,428	560,785	870,710	2,507,193	2,845,803	4,736,202	7,545,287	4,759,187	850,236	45,124
400	371,663	544,333	856,690	2,488,278	2,831,205	4,696,598	7,577,037	4,740,147	855,286	44,918
400	378,968	543,235	863,800	2,480,970	2,860,776	4,745,724	7,657,111	4,733,220	844,191	43,971
400	370,162	546,627	881,689	2,511,807	2,839,993	4,780,048	7,643,978	4,713,466	853,200	42,811
1000	1,020,528	1,464,705	2,185,549	5,174,790	5,663,302	9,564,030	14,913,140	9,911,388	1,539,691	117,387
1000	1,039,694	1,495,585	2,229,515	5,194,331	5,712,828	9,668,085	15,121,333	10,046,133	1,550,220	116,105
1000	1,015,193	1,460,448	2,192,634	5,215,365	5,824,863	9,789,186	15,143,680	9,370,015	1,566,303	119,394
1000	988,377	1,403,698	2,190,586	5,285,734	5,791,527	9,705,058	15,195,246	9,932,084	1,553,824	121,236
1500	1,448,538	2,031,036	3,056,378	6,953,608	7,583,410	12,641,907	19,335,848	12,327,199	1,908,636	174,429
1500	1,425,929	2,053,094	3,019,015	6,981,940	7,563,694	12,518,071	19,172,897	12,752,422	1,888,935	171,710
1500	1,469,290	2,033,443	3,030,611	6,904,557	7,584,307	12,626,396	19,308,182	12,842,599	1,902,875	175,247
1500	1,470,121	2,062,210	3,054,469	6,940,716	7,597,353	12,608,467	19,417,533	12,906,640	1,924,470	174,086

[22] [23], provided that the concentrations are sufficiently high. The PFOA shows an abnormal behaviour in this sense, since its CV first decreases and then increases.

It has been tried to establish a linear range of work in a smaller range of concentrations, eliminating for that in the calibration curve the points placed to the concentration 1000 and 1500 ppb (Figures 9-11). Although the R^2 values thus




Figure 9. Calibration curve obtained by simple linear regression eliminating the points of 1000 and 1500 ppb (top) and residual graph (bottom) for MeP, EtP, PrP and PFBuA.




Figure 10. Calibration curve obtained by simple linear regression eliminating the points of 1000 and 1500 ppb (top) and residual graph (bottom) for PFPeA, PFHxA, PFHpA and PFOA.




Figure 11. Calibration curve obtained by simple linear regression eliminating the points of 1000 and 1500 ppb (top) and residual graph (bottom) for PFOS and HBCDD.

obtained are greater than 0.99 in most cases, the residuals show in this case an upward or downward trend. In cases where the curvature is apparent, a quadratic equation model (second degree polynomial) to the data (Figure 12), obtaining a considerable improvement in the values of R^2 , being these of the order of 0.999, being the residuals above and below the zero, but not in a typical random pattern. This situation is not corrected with higher polynomial models, or with rational models [24] [41] of the type

$$y = \frac{a_0 + a_1x + a_2x^2}{1 + a_3x + a_4x^2} \quad (1)$$

As stated by Box: “There are no perfect models, but models that fit better than others” [42] [43] [44]. Linear or quadratic models, simpler, allow the calculation of concentrations with the required accuracy at the level of ppb, in which we are involved. The search of possible causes due to this phenomenon, as well as weighting factors to apply in the calibration and an analysis of the data in depth will be object of further search.

4. Final Comments

Calibration is an essential part of every quantitative analytical method and correct performance of the so important step is a critical part of method development and validation. Analytical chemists are often interested in the fitting of mathematical equations to experimental data [6] [8] [17] [18] [21] [22] [23].

The least squares method is widely used to find or estimate the numerical values of the parameters to fit a function to a set of data and to characterize the statistical properties of estimates. In spite of this, common situations when working with LC-MS/MS or absorption spectrophotometry that may be described by functional relationships include calibration curves relating measured values of response to a property, which may be nonlinear [4] [8] [11] [12] [24] [41].

In most of situations, a statistical test for linearity between the variables is rarely undertaken in analytical studies despite the frequent assumption that such linearity prevails. Taylor and Schutsyner [45] quoted in 1986: “*Although the theory concerning regression has since long been described, many errors can still be*




Figure 12. Calibration curve obtained by quadratic adjustment (simple linear regression) (top) and residual graph (bottom) for the studied compounds.

encountered when it is applied to solve problems in analytical chemistry.”

Residual analysis is a very useful tool that helps select the model that fit more adequately the data. In considering residuals, a qualitative approach is often the most revealing and informative.

In many chemical, pharmaceutical and biological applications the use of LC and MS have proved to be a powerful tool for the identification and quantification of multiresidue compounds in complex mixtures and numerous new methods are developed and validated daily. So that, a preliminary study on the simultaneous determination of compounds of environmental significance (six-perfluoroalkyl compounds, three preservatives and a brominated flame retardant) by LC-MS/MS has been carried out in this work. For calibration purposes, nine concentration levels were prepared and calibration curve was built. Most of the studied compounds show curvilinear calibration curves, which is not surprising given the wide concentrations range used. By restricting the concentra-

tion range a linear region may be sometimes choice to determine some of the compounds of environmental concern subject to study in this paper. By using parabolic regression, the dynamic range of some of the standard curves may be broader.

Note that the best choice from a practical point of view is the simplest model, which fit properly the data, in agreement with the parsimony principle (Occam's razor) [25]. However, things are no easy. As stated by Box [42] [43] [44] "all models are wrong". There are no perfect models, but model that are more adequate than others.

References

- [1] Eurachem Working Group (1998) Eurachem Guide: The Fitness for Purpose of Analytical Methods—A Laboratory Guide to Method Validation and Related Topics. Teddington, 3, 20-21.
- [2] Harmonized IUPAC Guidelines (2003) Guidelines for Single-Laboratory Validation of Methods of Analysis. CAC/GL 49-2003, Cambridge, 839, 846-847.
- [3] Sanagi, M.M., Nasir, Z., Ling, S.L., Hermawan, D., Ibrahim, W.A.W. and Naim, A.A. (2010) A Practical Approach for Linearity Assessment of Calibration Curves Under the International Union of Pure and Applied Chemistry (IUPAC) Guidelines for an In-House Validation of Method of Analysis. *Journal of AOAC International*, **93**, 1322-1330.
- [4] Asnin, L.D. (2016) Peak Measurement and Calibration in Chromatographic Analysis. *Trends in Analytical Chemistry*, **81**, 51-62. <https://doi.org/10.1016/j.trac.2016.01.006>
- [5] Kóscielniak, P. and Wiczorek, M. (2016) Univariate Analytical Calibration Methods and Procedures: A Review. *Analytica Chimica Acta*, **944**, 14-28. <https://doi.org/10.1016/j.aca.2016.09.024>
- [6] Raposo, F. (2016) Evaluation of Analytical Calibration Based on Least Squares Linear Regression for Instrumental Techniques: A Tutorial Review. *Trends in Analytical Chemistry*, **77**, 167-185. <https://doi.org/10.1016/j.trac.2015.12.006>
- [7] Olivieri, A.C. (2015) Practical Guidelines for Reporting Results in Simple and Multi-Component Analytical Calibration: A Tutorial. *Analytica Chimica Acta*, **868**, 10-22. <https://doi.org/10.1016/j.aca.2015.01.017>
- [8] Mermet, J.-M. (2010) Calibration in Atomic Spectrometry: A Tutorial Review Dealing with Quality Criteria, Weighting Procedures and Possible Curvatures. *Spectrochimica Acta B*, **65**, 509-523. <https://doi.org/10.1016/j.sab.2010.05.007>
- [9] Tellinghuisen, J. (2009) Least Squares in Calibration: Weights, Nonlinearity, and Other Nuisances. *Methods Enzymology*, **454**, 259-285. [https://doi.org/10.1016/S0076-6879\(08\)03810-X](https://doi.org/10.1016/S0076-6879(08)03810-X)
- [10] Switaj-Zawadha, A., Konieczka, P., Pryk, E. and Namiesnik, J. (2005) Calibration in Metrological Approach. *Analytical Letters*, **38**, 353-376. <https://doi.org/10.1081/AL-200043431>
- [11] Lavagnini, I. and Magno, F. (2007) A Statistical Overview on Univariate Calibration, Inverse Regression, and Detection Limits: Application to Gas Chromatography/Mass Spectrometry Technique. *Mass Spectrometry Reviews*, **26**, 1-18. <https://doi.org/10.1002/mas.20100>
- [12] Lavagnini, I., Magno, F., Seraglia, R. and Traldi, P. (2006) Quantitative Applications of Mass Spectrometry. Wiley, New York. <https://doi.org/10.1002/0470029021>

- [13] Chatterjee, S. and Hadi, A.S. (2012) Regression Analysis by Example. 5th Edition, Wiley, New York, 98.
- [14] Draper, N.R. and Smith, H. (1998) Applied Regression Analysis. 3th Edition, Wiley, New York. <https://doi.org/10.1002/9781118625590>
- [15] Cornish-Bowden, A. (2014) Analysis and Interpretation of Enzyme Kinetics Data. *Perspectives in Science*, **1**, 121-125. <https://doi.org/10.1016/j.pisc.2014.02.010>
- [16] Meloun, M. and Militky, J. (2011) Statistical Data Analysis, a Practical Guide. Woodhead Publishing, New Delhy.
- [17] Asuero, A.G., Sayago, A. and González, A.G. (2006) The Correlation Coefficient: An Overview. *Critical Reviews in Analytical Chemistry*, **36**, 41-59. <https://doi.org/10.1080/10408340500526766>
- [18] Sayago, A., Boccio, M. and Asuero, A.G. (2004) Fitting Straight Lines with Replicated Observations by Linear Regression: The Least Squares Postulates. *Critical Reviews in Analytical Chemistry*, **34**, 39-50. <https://doi.org/10.1080/10408340490273744>
- [19] Wisniak, J. and Polishuk, A. (1999) Analysis of Residuals—A Useful Tool for Phase Equilibrium Data Analysis. *Fluid Phase Equilibria*, **164**, 61-82. [https://doi.org/10.1016/S0378-3812\(99\)00246-0](https://doi.org/10.1016/S0378-3812(99)00246-0)
- [20] Noggle, J.H. (1933) Practical Curve Fitting and Data Analysis. Software and Self-Instructions for Scientists and Engineers, Prentice Hall, Englewood Cliffs.
- [21] Asuero, A.G. and González, A.G. (2007) Fitting Straight Lines with Replicated Observations by Linear Regression. III. Weighting Data. *Critical Reviews in Analytical Chemistry*, **37**, 143-172. <https://doi.org/10.1080/10408340701244615>
- [22] Asuero, A.G. and Martin, J. (2011) Fitting Straight Lines with Replicated Observations by Linear Regression. IV. Transforming Data. *Critical Reviews in Analytical Chemistry*, **41**, 36-69. <https://doi.org/10.1080/10408347.2010.523589>
- [23] Asuero, A.G. and González, A.G. (1989) Some Observations on Fitting a Straight Line to Data. *Microchemical Journal*, **40**, 216-225. [https://doi.org/10.1016/0026-265X\(89\)90073-8](https://doi.org/10.1016/0026-265X(89)90073-8)
- [24] Michalowska-Kazcmarczyk, A., Asuero, A.G., Martin, J., Alonso, E., Jurado, J.M. and Michalowski, T. (2014) A Uniform Nonlinearity Criteria for Rational Functions Applied to Calibration Curve and Standard Addition Methods. *Talanta*, **130**, 307-314. <https://doi.org/10.1016/j.talanta.2014.04.067>
- [25] Bates, D.M. and Watt, D.G. (2007) Nonlinear Regression Analysis and Its Applications. 2nd Edition, Wiley, New York, 1.
- [26] Martín, J., Santos, J.L., Aparicio, I. and Alonso, E. (2015) Determination of Hormones, a Plasticizer, Preservatives, Perfluoroalkylated Compounds, and a Flame Retardant in Water Samples by Ultrasound-Assisted Dispersive Liquid-Liquid Microextraction Based on the Solidification of a Floating Organic Drop. *Talanta*, **143**, 335-343. <https://doi.org/10.1016/j.talanta.2015.04.089>
- [27] Farré, M., Kantiani, L., Petrovic, M., Pérez, S. and Barceló, D. (2012) Achievements and Future Trends in the Analysis of Emerging Organic Contaminants in Environmental Samples by Mass Spectrometry and Bioanalytical Techniques. *Journal of Chromatography A*, **1259**, 86-99. <https://doi.org/10.1016/j.chroma.2012.07.024>
- [28] Banjac, Z., Ginebreda, A., Kuzmanovic, M., Marcé, R., Nadal, M., Riera, J.M. and Barceló, D. (2015) Emission Factor Estimation of ca. 160 Emerging Organic Microcontaminants by Inverse Modeling in a Mediterranean River Basin (Llobregat, NE Spain). *Science of the Total Environment*, **520**, 241-252. <https://doi.org/10.1016/j.scitotenv.2015.03.055>

- [29] Fromme, H., Becher, G., Hilger, B. and Völkel, W. (2016) Brominated Flame Retardants—Exposure and Risk Assessment for the General Population. *International Journal of Hygiene and Environmental Health*, **219**, 1-23. <https://doi.org/10.1016/j.ijheh.2015.08.004>
- [30] Allgood, J.M., Jimah, T., McClaskey, C.M., La Guardia, M.J., Hammel, S.C., Zeineddine, M.M., Tang, I.W., Runnerstrom, M.G. and Ogunseitán, O.A. (2017) Potential Human Exposure to Halogenated Flame-Retardants in Elevated Surface Dust and Floor Dust in an Academic Environment. *Environmental Research*, **153**, 55-62. <https://doi.org/10.1016/j.envres.2016.11.010>
- [31] Rivière, G., Sirot, V., Tard, A., Jean, J., Marchand, P., Veyrand, B., Le Bizec, B. and Leblanc, J.C. (2014) Food Risk Assessment for Perfluoroalkyl Acids and Brominated Flame Retardants in the French Population: Results from the Second French Total Diet Study. *Science of the Total Environment*, **491-492**, 176-183. <https://doi.org/10.1016/j.scitotenv.2014.01.104>
- [32] Hendriks, H.S. and Westerink, R.H.S. (2015) Neurotoxicity and Risk Assessment of Brominated and Alternative Flame Retardants. *Neurotoxicology and Teratology*, **52**, 248-269. <https://doi.org/10.1016/j.ntt.2015.09.002>
- [33] The European Parliament and the Council of the European Union (2008) Directive 2008/105/EC of the European Parliament and of the Council of 16 December 2008 on Environmental Quality Standards in the Field of Water Policy, Amending and Subsequently Repealing Council Directives 82/176/EEC, 83/513/EEC, 84/156/EEC, 84/491/EEC, 86/280/EEC and Amending Directive 2000/60/EC of the European Parliament and of the Council. *Official Journal of the European Union*, **348**, 84.
- [34] Corsini, E., Luebke, R.W., Germolec, D.R. and DeWitt, J.C. (2014) Perfluorinated Compounds: Emerging POPs with Potential Immunotoxicity. *Toxicological Letters*, **230**, 263-270. <https://doi.org/10.1016/j.toxlet.2014.01.038>
- [35] Hekster, F.M., Laane, R.W.P.M. and de Voogt, P. (2003) Environmental and Toxicity Effects of Perfluoroalkylated Substances. *Reviews of Environmental Contamination and Toxicology*, **179**, 99-121. https://doi.org/10.1007/0-387-21731-2_4
- [36] Domingo, J.L. (2011) Health Risks of Dietary Exposure to Perfluorinated Compounds. *Environment International*, **40**, 187-195. <https://doi.org/10.1016/j.envint.2011.08.001>
- [37] Soni, M.G., Carabin, I.G. and Burdock, G.A. (2005) Safety Assessment of Esters of *p*-Hydroxybenzoic Acid (Parabens). *Food and Chemical Toxicology*, **43**, 985-1015. <https://doi.org/10.1016/j.fct.2005.01.020>
- [38] Boberg, J., Taxvig, C., Christiansen, S. and Hass, U. (2010) Possible Endocrine Disrupting Effects of Parabens and Their Metabolites. *Reproductive Toxicology*, **30**, 301-312. <https://doi.org/10.1016/j.reprotox.2010.03.011>
- [39] European Parliament and of the Council of 30 November 2009 on Cosmetic Products; European Union Regulation No. 1223/2009; 2009; 59-209.
- [40] Commission Decision (2002/657/EC) of 12 August 2002 Implementing Council Directive 96/23/EC Concerning the Performance of Analytical Methods and the Interpretation of Results, Official Journal of the European Communities L221, Brussels, Belgium, 8-36.
- [41] Gorazda, K., Michalowska-Kaczmarczyk, A.M., Asuero, A.G. and Michalowski, T. (2013) Application of Rational Functions for the Standard Addition Method. *Talanta*, **116**, 927-930. <https://doi.org/10.1016/j.talanta.2013.07.085>
- [42] Box, G.E.P. (1976) Science and Statistics. *Journal of the American Statistical Association*, **71**, 791-796. <https://doi.org/10.1080/01621459.1976.10480949>
- [43] Box, G.E.P. and Draper, N.R. (1987) Empirical Model Building and Response Sur-

faces. Wiley, New York.

- [44] Box, G.E.P., Hunter, J.S. and Hunter, W.G. (2005) *Statistics for Experimenters*. 2nd Edition, Wiley, New York.
- [45] Taylor, P.D.P. and Schutyser, P. (1986) Weighted Linear-Regression Applied in Inductively Couple Plasma-Atomic Emission Spectrometry—A Review of the Statistical Considerations Involved. *Spectrochimica Acta*, **41B**, 1051-1061.



Scientific Research Publishing

Submit or recommend next manuscript to SCIRP and we will provide best service for you:

Accepting pre-submission inquiries through Email, Facebook, LinkedIn, Twitter, etc.

A wide selection of journals (inclusive of 9 subjects, more than 200 journals)

Providing 24-hour high-quality service

User-friendly online submission system

Fair and swift peer-review system

Efficient typesetting and proofreading procedure

Display of the result of downloads and visits, as well as the number of cited articles

Maximum dissemination of your research work

Submit your manuscript at: <http://papersubmission.scirp.org/>

Or contact jasmi@scirp.org

